



Alan C. Lloyd, Ph.D.
Agency Secretary

Air Resources Board

Robert F. Sawyer, Ph.D., Chair
1001 I Street • P.O. Box 2815
Sacramento, California 95812 • www.arb.ca.gov



Arnold Schwarzenegger
Governor

August 8, 2008

Mr. Michael Lessley
Triangle Environmental, Incorporated
730 North Mariposa Street
Burbank, California 91506-1629

Dear Mr. Lessley:

Thank you for your letter dated May 30, 2008 requesting Air Resources Board's (ARB) approval of the TriTester Model TVLM-00 (TriTester), Version DR software to be used to conduct Exhibit 11 of Executive Order (EO) VR-204-C. The TriTester is currently equipped with Version 2.96 and Version 2.01 which were deemed equivalent to ARB's TP-201.5 (1996 version) and Exhibit 5 of EO VR-201 series and VR-202 series, respectively. Version DR software will be added to the Tritester so that it can also be used to conduct Exhibit 11 of EO VR-204-C.

The requirements for a gas volume meter outlined in Exhibit 11 of EO VR-204-C state that: "...Use a Dresser Measurement Roots Meter®, or equivalent (preferably fitted with a digital readout)...". The TriTester with Version DR software does not require ARB approval since it meets this requirement. Therefore, the TriTester with Version DR software can be used to conduct Exhibit 11 of EO VR-204-C.

If you have questions or need further information, please contact either Sam Vogt at (916) 322-8922 or via email at svogt@arb.ca.gov.

Sincerely,

Manjit Ahuja for WVL

William V. Loscutoff, Chief
Monitoring and Laboratory Division

cc: See Next Page

Mr. Michael Lessley
August 8, 2008
Page 2

Enclosure

cc: Brian Aungar
San Luis Obispo Air Pollution Control District

Jim Swaney
San Joaquin Valley Air Pollution Control District

Jeannette Lim
Bay Area Air Quality Management District

Randy Matsuyama
South Coast Air Quality Management District